


CERTIFICATE OF ANALYSIS

KDM5D Probe Set (Orange; 20 Test Kit; 40 µl)

Catalog # KDM5D-20-O-R



| Fluorochrome | Localization |
|--------------|--------------|
| Orange | Yq11.222 |

Lot Number: 2025-03-26

Expiration Date: March 2027

Product Description – Quality Declaration

The **KDM5D Probe Set** consists of DNA labeled in Spectrum Orange. The DNA probe set hybridizes to chromosome Yq11.222 in normal metaphase spreads and interphase nuclei.

Disclaimer

This probe set is for RESEARCH USE ONLY.

The probe set has NOT BEEN TESTED ON CLINICAL SAMPLES and is NOT VALIDATED.

Performance and analytical characteristics HAVE NOT been determined.

Chromosomal location based on GRCh37/hg19 build of human genome assembly.

Third party expiration studies were completed on concentrated probes only. We cannot guarantee the expiration date of probes mixed with buffer and stored reconstituted.