


CERTIFICATE OF ANALYSIS

MYC (RTU) Probe Set (Red; 50 Test Kit; 500 µl)

Catalog # MYC-RTU-50-RE



Fluorochrome	Localization
Red	8q24.21

Lot Number: 2023-05-10

Expiration Date: May 2026

Product Description – Quality Declaration

The **MYC (RTU) Probe Set** consists of DNA labeled in Spectrum Red. The DNA probe set hybridizes to chromosome 8q24.21 in normal metaphase spreads and interphase nuclei.

Disclaimer

This probe set is for RESEARCH USE ONLY.

The probe set has NOT BEEN TESTED ON CLINICAL SAMPLES and is NOT VALIDATED.

Performance and analytical characteristics HAVE NOT been determined.

Chromosomal location based on GRCh37/hg19 build of human genome assembly.

Third party expiration studies were completed on concentrated probes only. We cannot guarantee the expiration date of probes mixed with buffer and stored reconstituted.