


CERTIFICATE OF ANALYSIS

RET BA Probe Set (Orange/Green; 20 Test Kit; 40 µl)

Catalog # RETBA-20-ORGR



Fluorochrome	Localization
Orange	10q11.21
Green	10q11.21

Lot Number: 2022-06-16

Expiration Date: 2025-06-15

Product Description – Quality Declaration

The **RET BA Probe Set** consists of DNA labeled in Spectrum Orange and Spectrum Green. The DNA probe set hybridizes to chromosome 10q11.21 in normal metaphase spreads and interphase nuclei.

Disclaimer Product Description – Quality Declaration

This probe set is CE MARKED in certain countries.

This probe set is for RESEARCH USE ONLY elsewhere.

The probe set has NOT BEEN TESTED ON CLINICAL SAMPLES and is NOT VALIDATED.

Performance and analytical characteristics HAVE NOT been determined.

Chromosomal location based on GRCh37/hg19 build of human genome assembly.

Third party expiration studies were completed on concentrated probes only. We cannot guarantee the expiration date of probes mixed with buffer and stored reconstituted.

EC REP MedNet EC-REP GmbH Borkstrasse 10 48163 Münster Germany

