

# CERTIFICATE OF ANALYSIS

TP53/Con 17 FISH Probe (Orange/Green; 20 Test Kit; 40 µl)

Catalog # TP53-CHR17-20-ORGR

## TP53/Control 17 (17p13.1/17p11.1)




IMAGE NOT DRAWN TO SCALE

Fluorochrome	Localization
Orange	17p13.1
Green	17p11.1

Lot Number: 2022-05-11

Expiration Date: 2025-05-05

### Product Description – Quality Declaration

The **TP53/Con 17 FISH Probe** consists of DNA labeled in Spectrum Orange and Spectrum Green. The DNA probe set hybridizes to chromosome 17p13.1 (Orange) and 17p11.1 (Green) in normal metaphase spreads and interphase nuclei.

### Disclaimer

This probe set is CE MARKED in certain countries.

This probe set is for RESEARCH USE ONLY elsewhere.

The probe set has NOT BEEN TESTED ON CLINICAL SAMPLES and is NOT VALIDATED.

Performance and analytical characteristics HAVE NOT been determined.

Chromosomal location based on GRCh37/hg19 build of human genome assembly.

\*\*\*Third party expiration studies were completed on concentrated probes only. We cannot guarantee the expiration date of probes mixed with buffer and stored reconstituted.\*\*\*

EC REP MedNet EC-REP GmbH Borkstrasse 10 48163 Münster Germany